

Expert Advice

v2.0



Malvern Instruments Ltd - © Copyright 2008

Size measurements

Insufficient signal collected (7048 < 10000 k total photons).
Measurement duration not set to automatic.
Filter factor in research software not set to default (50%).